

CNC and Polaris Comparison Test Results - March 8th, 2012

/\*-----\*/

P\_X CNC\_Z 140mm

/\*-----\*/

Maximum Error: 0.15 mm  
Mean Error: 0.04 mm  
Standard Deviation: 0.03  
Coefficient of Variation: 0.73

/\*-----\*/

P\_Y CNC\_X 400mm

/\*-----\*/

Maximum Error: 0.18 mm  
Mean Error: 0.07 mm  
Standard Deviation: 0.05  
Coefficient of Variation: 0.70

/\*-----\*/

P\_Z CNC\_Y 200mm

/\*-----\*/

Maximum Error: 0.19 mm  
Mean Error: 0.07 mm  
Standard Deviation: 0.05  
Coefficient of Variation: 0.68